

Feasibility of the β^- Radio-Guided Surgery with a Variety of Radio-Nuclides of Interest to Nuclear Medicine

Carlo Mancini-Terracciano^{1,2}, Raffaella Donnarumma*^{2,3},
Gaia Bencivenga⁴, Valerio Bocci², Antonella Cartoni⁵,
Francesco Collamati^{1,2}, Ilaria Fratoddi⁵, Alessandro Giordano⁶,
Luca Indovina⁷, Michela Marafini⁸, Silvio Morganti²,
Dante Rotili⁹, Andrea Russomando^{2,3,10}, Teresa Scotognella¹¹,
Elena Solfaroli Camillocci^{2,3}, Marco Toppi¹², Giacomo Traini^{2,3},
Iole Venditti⁵, and Riccardo Faccini^{2,3}

¹Dip. Scienze di Base e Applicate per l'Ingegneria, Sapienza Univ. di Roma, Rome, Italy;

²INFN Sezione di Roma, Rome, Italy;

³Dip. Fisica, Sapienza Univ. di Roma, Rome, Italy;

⁴PET-CT Center, Policlinico A. Gemelli, Rome;

⁵Dip. Chimica, Sapienza Univ. di Roma, Rome, Italy;

⁶Ist. Medicina Nucleare, Univ. Cattolica del Sacro Cuore, Rome, Italy;

⁷UOC Fisica Sanitaria, Policlinico A. Gemelli, Rome;

⁸Museo Storico della Fisica e Centro Studi e Ricerche "E. Fermi", Rome, Italy;

⁹Dip. Chimica e Tecnologie del Farmaco, Sapienza Univ. di Roma, Rome, Italy;

¹⁰Center for Life Nano Science@Sapienza, Istituto Italiano di Tecnologia, Rome, Italy.

¹¹Ist. Medicina Nucleare, Policlinico A. Gemelli, Rome;

¹²Laboratori Nazionali di Frascati INFN, Frascati, Italy;

October 31, 2016

Abstract

The β^- based radio-guided surgery overcomes the corresponding γ technique in case the background from healthy tissues is relevant. It can

*co-first author

be used only in case a radio-tracer marked with ^{90}Y is available since the current probe prototype was optimized for the emission spectrum of this radio-nuclide. Here we study, with a set of laboratory tests and simulations, the prototype capability in case a different radio-nuclide is chosen among those used in nuclear medicine.

As a result we estimate the probe efficiency on electrons and photons as a function of energy and we evaluate the feasibility of a radio-guided surgery exploiting the selected radio-nuclides. We conclude that requiring a 0.1 ml residue to be detected within 1 s by administering 3 MBq/Kg of radio-isotope, the current probe prototype would yield a significant signal in a vast range of values of SUV and TNR in case ^{31}Si , ^{32}P , ^{97}Zr , and ^{188}Re are used. Conversely, a tuning of the detector would be needed to efficiency use ^{83}Br , ^{133}I , and ^{153}Sm , although they could already be used in case of high SUV or TNR values. Finally, ^{18}F , ^{67}Cu , ^{131}I , and ^{177}Lu are not useable for radio-guided surgery with the current probe design.

Introduction

Radio-guided surgery (RGS) is a technique that enables the surgeon to evaluate in real time the completeness of the tumor lesion resection. At the same time RGS allows to minimise the amount of healthy tissue removed. It represents a significant adjunct to intra-operative detection of millimetre-sized tumor residues, providing the surgeon with vital and real-time information regarding the location and the extent of the lesion. It is crucial for those tumors where surgical resection is the only possible therapy, since it reduces the probability of tumor recurrence assessing surgical resection margins.

RGS makes use of a radio-labelled tracer and of a probe. The radiopharmaceutical is administered to the patient before surgery and its uptake has to be higher in the tumor than in healthy organs, allowing to discriminate among them.

Traditional RGS makes use of γ emitters as tracer [1, 2, 3]. However, the use of this technique is strongly limited, and often even prevented, by the background produced from the uptakes of the healthy tissue. Such a background is non-negligible because γ radiation can travel through large amounts of tissue. Moreover, the large mean free path of γ radiation exposes the medical personnel to radiation risks.

The use of β^- radiation to overcome the limitations of traditional RGS was recently proposed [4, 5]. Indeed, electrons have a very small penetration. Therefore a β^- RGS has a more favourable ratio between the signal coming from the tumor and the rest of the body. An ex-vivo test has shown the feasibility of this technique for a meningioma case [6]. Meningioma has been chosen because it is known to express receptors to DOTATOC [7], a somatostatin analogue that could be labelled with ^{90}Y , a pure β^- emitter with a 2.3 MeV electron energy endpoint.

Limiting to such radio-tracer, though, would mean restricting this technique to very few tumors, mainly glioma [8] and neuro-endocrine tumors [9]. To further broaden the applicability of the technique, the use of different radio-isotopes, with larger gamma contamination and lower endpoint is under investigation.

Among the β^- -emitting radio-isotopes that have the largest diffusion in nu-

Isotope	$\tau_{1/2}(h)$	E_γ (keV)	I_γ (%)	EP_β (keV)	I_β (%)
^{31}Si	2.62			1491	100
^{32}P	343			1710	100
^{67}Cu	62	93 184	16 48	377	57
				468	22
				561	20
^{83}Br	2.4			935	99
^{90}Y	64			2280	100
^{97}Zr	17	743	93	759	88
(secondary*)		657	86	1277	86
^{131}I	192	365	82	334	7
		637	7	606	90
^{133}I	20.8	530	87	1227	83.4
^{153}Sm	46	103	29	635	31
				704	49
				808	18
^{177}Lu	160	112	6	177	12
		208	10	500	79
^{188}Re	17	155	15	1962	25
				2118	72

Table 1: Characteristics of the β^- radio-isotopes of interest for the proposed RGS technique: half-life ($\tau_{1/2}$), γ and β^- intensities (I_γ , I_β), photon lines energy (E_γ), and β^- end-points (EP_β). Only components with $I > 5\%$ are listed. (*) the second line of ^{97}Zr is due to the subsequent decay, with $\tau_{1/2} = 72$ min of the ^{97}Nb produced in the primary decay.

clear therapy [10], we considered those that have an electron energy end-point of at least 500 keV, dominant γ lines below 300 keV and total γ intensities below 100%: ^{32}P , ^{131}I , ^{133}I , ^{153}Sm , ^{177}Lu , and ^{188}Re . In addition we considered isotopes that are chemically in the same family as nuclides already used to mark tracers since these are the best candidates to give life to a new radio-tracer. We thus included in our studies ^{31}Si , ^{67}Cu , ^{83}Br , and ^{97}Zr . The characteristics of the decays of the considered isotopes are detailed in Tab. 1.




Figure 1: The probe prototype and the experimental setup of the measurements with the sealed point-like sources (left) and scheme of the setup with ^{18}F (right).

Another alternative is to consider ^{18}F , which has $\tau_{1/2} = 1.8$ h and a β^+ emission with end-point at 633.5 keV and 97% intensity and is used to radio-label the fluorodeoxyglucose (FDG). The uptake of the latter is the same as glucose, which in turn is correlated with tissue metabolism. It is widely used as marker for the positron emission tomography (PET) and, therefore, if the developed probe were sensitive to ^{18}F with low background there would be a significant

extension of the applicability of the technique. It is to be noted that in this case the probe would detect the positrons before they stop or annihilate, while annihilation photons would represent a background. The possibility to intra-operatively detect β^+ decays has already been explored in the past [11, 12] with the development of ad-hoc dual detectors for the simultaneous measurement of positrons and γ s. The difficulties in developing and handling the probe has limited the clinical studies [13].

This paper describes the laboratory tests performed to measure the probe efficiency to γ and β radiations, studies that are the basis for the extension of the technique to radio-isotopes other than the baseline ^{90}Y . To investigate potentialities of radiations significantly different, several sources with different β spectra and gamma lines were used.

After tuning the simulation on these measurements, the expected signal from tumor and background assuming to use the different possible radionuclides of Tab. 1 was estimated. Finally in the comparison among radio-isotopes we also estimated the expected dose rate delivered to the medical personnel during surgery. This requirement, in addition to the need to minimize the dose to the patient, limits the activity that can be administered.

Material and Methods

The Probe Prototype

The probe prototype considered in this paper is visible in Fig. 1. The lower black part in the picture is a ring of 12 mm external diameter in Acrylonitrile Butadiene Styrene (ABS) that shields the sensitive element content in it from the radiation coming from the side. The sensitive part of the probe is a 3 mm height and 5 mm diameter of mono-crystalline para-terphenyl doped to 0.1% in mass with diphenylbutadiene. This plastic scintillator was adopted because of its high light yield (~ 3 times larger than typical organic scintillators), non-hygroscopic property, and low density [14] that minimises the sensitivity to photons. The light tightness around the scintillator is ensured by a 15 μm thick aluminum sheet. The scintillation light is read out by a SiPM (sensL B-series 10035) biased with 24.5 V. Its spectral range is 300 – 800 nm and the peak wavelength is 420 nm. An aluminum cylindrical body (diameter 12 mm and length 14 cm) encapsulates this assembly. A portable electronics based on ArduSiPM, with ethernet connection to a PC was used for the read out [15].

Experimental Setups

We measured the probe sensitivity to the radiation emitted by several long life-time radioisotopes with different energies, contained in sealed sources and in liquid ^{18}F samples. They contain radio-isotopes by definition different from the ones of interest to RGS, which requires half-lives shorter than few days. Therefore these measurements can only be used to optimize the parameters used in the MC simulations (see Sec.) to describe the probe behavior, in particular the optical properties of para-terphenyl. The MC is in turn used to estimate the RGS performances.

This measurement campaign has been performed with two different setups:

the setup used for the sealed laboratory sources and the one made to measure the probe efficiency to liquid ^{18}F . These two setups will be described separately in the next subsections.

Isotope	A (kBq)	$\tau_{1/2}$ (y)	E_γ (keV)	I (%)	$EP_{\beta-}$ (keV)	I (%)	$EP_{\beta+}$ (keV)	I (%)	e^- (keV)	I (%)
^{60}Co	36.7 ± 1.1	5.3	1173 1332	100 100	317	100				
^{90}Sr	3.3 ± 0.1	28.9			540 2280	100 100				
^{22}Na	40.5 ± 1.2	2.6	511 1274	181 100			546	90		
^{152}Eu	41.5 ± 1.2	13.2	40 45 28 244 344 443 779 867 964 1085 1112 1408	58 11 1475 8 27 3 13 4 15 12 4 21	384 694 8	2.4 14			75 114 2.4	19 11
		122						120		

Table 2: Characteristics of the sealed sources used in the tests: activity at the moment of the experiment (A), half-life ($\tau_{1/2}$), intensities (I), photon lines energy (E_γ), and end-points (EP).

Sealed sources setup

The sealed sources were chosen in order to test the efficiency to photons and electrons at several energies. Their relevant characteristics are listed in Tab. 2. The γ sources have also β decays from internal conversion. Since the amount of electrons exiting the source depends strongly on the its details, we added a 2.8 mm copper shielding.

All the sealed sources were point-like a part the ^{90}Sr one that had a diameter of 16 mm . The probe was placed vertically on top of each of them, 3.5 mm far from the sources. A picture of the experimental setup is in the left of Fig. 1. A bi-dimensional horizontal scan was performed for each measurement in order to center the probe with respect to the source.

Finally, to study different electron spectra, the measurements with the ^{90}Sr were repeated also with a plastic shielding made of up to 11 sheets of 2-methyl-1,3-butadiene, 400 μm thick.

^{18}F setup

To measure the probe efficiency to ^{18}F , 0.48 ml of water with 53 kBq of such isotope have been placed in a plastic (PVC) vessel, its central water receptacle was cylindrically shaped, with a diameter of 12.2 mm. To shield from γ radiation, the plastic vessel has been placed in a lead vial, also cylindrically shaped. The lead vial was much larger than the plastic vessel to reduce the background due to the annihilation γ 's. In detail: the vial internal radius was 2 cm and the distance from the bottom 4.4 cm. A scheme of the experimental setup, with the most relevant sizes of the lead vial shield and the vessel, is shown in Fig. 1 on the left.

The Probe Simulation and its Optimization

The estimate of the expected performances of the β^- probe prototype in case of use of β^+ radio-tracers for intra-operatorial use requires an estimate of the signal from the tumor and of the background coming from the uptake of radio-tracer from the rest of the body. These measurements need a simulation properly reproducing the measurements on the sources.

Furthermore, this study aims to quantify the efficiency on electrons and photons as a function of their energies and this also requires a detailed simulation.

To this aim, the FLUKA Monte Carlo (MC) code [16] was used for a full simulation of the probe and of the experimental setups. From the geometry of the setup and of the probe the program estimates the energy released in the interaction of the decay products of the radio-nuclides with the detector and produces 28000 optical photons per MeV of deposited energy [14]. To increase the simulation precision, the δ rays production threshold and the electron transport threshold have been set to 10 keV in the scintillator and in the material in contact with it. Moreover, in these materials the electrons step size has been limited to make them lose no more than 1% of their energy per step. This allows to accurately reproduce the optical photons generation points and the δ rays production at the interface with the scintillator. In all the other materials the transport threshold of all the particles and the δ ray production threshold are set at 100 keV. The path of the optical photons is traced within the probe and the number of optical photons reaching the light detector (SiPM) is counted (N_{op}). The fraction of the decaying nuclei that release a signal is computed as the fraction of simulated decays that has N_{op} greater than a threshold (N_{th}) representing the minimum signal detectable by the probe prototype. Such efficiency (ϵ^{MC}) depends therefore on this threshold.

The simulation has several parameters which need to be set empirically. The ones that are most relevant for this analysis are related to the optical properties of the para-terphenyl, since they determine the conversion between the released energy and N_{op} . While the attenuation length has been measured to be ≈ 2 cm, the diffusion coefficient (D expressed in cm^{-1}) and the reflectivity (r) of the aluminum-wrapping were varied to find the configuration that best reproduces the measurements. Finally also N_{th} , related to the electronic threshold applied at readout, is optimized on data.

The optimization of these parameters is based on the measurement of the maximum observed rates in the scans over $N_s - 1$ sealed sources described in Sec. ¹ and the ^{18}F setup described in Sec. : R_i , $i = 1, \dots, N_s$. Such measurements are divided by the corresponding source activity (A_i), to estimate the efficiency on data, $\epsilon_i^{DT} = R_i/A_i$, with an error estimated as the sum in quadrature of the statistical error on R_i and the uncertainty on A_i (see Tab. 2).

At the same time, for each source the MC estimates the efficiency for each possible value of the free parameters ($\vec{P} = \{D, r, N_{th}\}$): $\epsilon_i^{MC}(\vec{P})$. The best estimates of the parameters are those that minimize

$$\chi^2(\vec{P}) = \sum_i \left(\frac{\epsilon_i^{MC}(\vec{P}) - \epsilon_i^{DT}}{\sigma_i} \right)^2. \quad (1)$$

¹of the described setups we used in the minimization only part of the measurements with ^{90}Sr , both to avoid giving too much weight on the electron detection and to have a control sample (see right of Fig. 3).

Estimate of expected RGS performances

In order to quantify the performances of the probe on different radio-isotopes, a benchmark surgical configuration was established. We considered that a typical patient of 70 kg of weight could be administered ≈ 210 MBq, and that a typical SUV of interest could range between 1 and 8, with a ratio between the uptake in the tumor and in the nearby healthy tissue (TNR) ranging between 4 and 30, as established in our previous studies on cerebral tumors [8].




Figure 2: Geometry of the simulation used to estimate the far background

To estimate the discrimination power between tumor and healthy tissues, we performed, for each radio-isotope, two sets of MC simulations: one to evaluate the signal from a tumor residue and the background due to the radiation emitted from the nearby healthy tissue; the other one to evaluate the background, mostly due to γ emission resulting from the uptake in the whole body. The acquired information was then combined to estimate the minimum time needed to identify a tumor residue or the minimum specific activity that needs to be administered to have a timely response from the probe.

Signal From Tumor and Nearby Healthy Tissues

To estimate the signal that we would measure from a small residue embedded in healthy tissue, we simulated a 6 mm diameter, 7 mm height cylinder with activity $A_T = A_M * SUV$ (where A_M is the mass specific administered activity), representing the residue, encapsulated in the center of a 2 cm diameter, 1 cm height cylinder with activity $A_B = A_T / TNR$, representing the healthy tissue. The estimated rate is R_{Tum} .

Conversely, to estimate the signal we would have on the nearby healthy tissue, a 2 cm diameter, 1 cm height cylinder completely active with A_B was simulated. The estimated rate is R_{Near} .

Far Background

The simulation was performed on a sample PET/CT scan provided by the Osirix web site² (see Fig. 2). The CT voxels, have been converted from Hounsfield Units (HU) into density and elemental composition for MC simulation[17]. The spatial distribution of the isotopes is sampled using the FDG-PET scan of the same patient, thus assuming for the hypothetical radio-tracer the same uptake as the FDG.

² the DICOM file called “MELANIX” on <http://www.osirix-viewer.com/datasets/>

To increase the simulation efficiency, and to average on different possible positions of the probe with respect to the patient body, 50 replicas of the probe were simulated and the average counting rate (R_{Far}), scaled for the total activity in the patient, was considered as background rate. The probes were conservatively placed on the abdomen of the patient.

To avoid double counting the background from nearby healthy tissues, in the estimate of the far background contributions from decays in a fiducial volume around the probe were discarded. Eventually, the signal expected when the probe is in contact with a tumor residue (S_{res}) or healthy tissue (S_{HT}) are estimated as:

$$S_{Res} = R_{Tum} + R_{Far} \quad (2)$$

$$S_{HT} = R_{Near} + R_{Far} \quad (3)$$

Minimum Time Needed and Administered Activity

The benchmark of the feasibility of the use of a given radiation is the activity that needs to be administered to have a significant signal in a reasonable time. We consider to be able to efficiently detect residues if the probability of false negatives $FN < 5\%$ and of false positives $FP < 1\%$ ³. Under these conditions, as described in Ref. [8] we can calculate t_{probe}^{min} , the time needed to have such a response given the administered activity and A_{1s}^{min} , the mass specific activity needed to have a reasonable time response of $t_{probe}^{min} = 1$ s.

Estimate of Medical Staff Exposure

Another critical information to establish the feasibility of an RGS technique is the dose received by the medical personnel. Given that a reasonable limit on the dose received by the medical personnel is 1 mSv/year, the dose rate induced by the activity administered to the patient defines the maximum number of hours that the surgical staff can operate adopting the RGS.

Given the limited penetration of the β^- radiation we consider that the dominant contribution to the dose to the surgical staff comes photons. Therefore, we estimate the expected effective dose rate from the specific gamma-ray dose rate (Γ) calculations in Ref. [18]. In the case of pure β^- decays we consider the effective dose delivered by the Brehmsstrahlung radiation as computed in Ref. [19]. From the estimated Γ values we extract DR_{med} , the expected dose rate on the medical staff in case 3MBq/kg of radio-isotope is administered to a 70 kg patient and the person is on average at a distance of 1 m. The estimated values are reported in Tab. 3.

Results

Simulation Parameter Optimization

The optimization of the MC parameters, the diffusion coefficient D , the reflectivity (r) and the equivalent threshold in optical photons (N_{th}), were determined as described in Sec. .

³FN and FP are computed from S_{Res} and S_{HT} as described in Ref. [8]




Figure 3: Left: χ^2 as a function of the diffusion coefficient (D) and the reflectivity (r) after the optimization of N_{th} . Right: Comparison between experimental counts and MC for the optimized values of D and r . The residual discrepancy (exp/MC) is also shown. Green bins in the bottom panel indicate the sources included in the χ^2 minimization.

The measured χ^2 as a function of D and r is shown in Fig. 3 on the left. There is a clear correlation between the two parameters, a lower reflectivity being compensated by a lower diffusion parameter. Nonetheless there is a well defined minimum in $r=80\%$ and $D=1.2\text{ cm}^{-1}$ and these are the adopted parameters in the remaining of the paper.

Fig. 3 on the right shows the resulting level of agreement of the measured and estimated rates together with the fractional discrepancy between the two.

Efficiencies

After fixing the MC parameters, the efficiency of the probe on electrons and photons could be estimated as a function of the energy. The result is reported in Fig. 4: the efficiency on electrons reaches a plateau of 95%, the flexus of the distribution being $\approx 400\text{keV}$. Similarly, the photon efficiency reaches 1.2% at $\approx 800\text{ keV}$ and has a less steep raise.

These observations confirm that the probe is optimal for ^{90}Y and that it has low sensitivity to photons. The performances are also significantly better than the previous prototype with PMT readout [20].




Figure 4: Probe efficiency as a function of the electron (top) and photon (bottom) energy.

RGS Feasibility

In order to compare the expected performances of the current probe prototype among radio-isotope, the simulation has been run as described in Sec. , assuming a 70 kg patient with SUV ranging between 1 and 8 and TNR between 4 and 30. Tab. 3 reports the relevant information for a particular case, with a reasonable SUV (4) and TNR (8). As a reference also the results obtained by the same analysis with the golden radio-isotope, ^{90}Y , are reported.

The extrapolation to other values of SUV and TNR are shown in Fig. 5.

Table 3: Comparison between radio-isotopes: the expected signal on a residue (S_{res}) and on healthy tissues (S_{HT}) and the contribution from the far background (R_{Far}); the minimum time required to identify a residue as defined in the text (t_{probe}^{min}); the dose rate to the medical personnel (DR_{med}) if 3 MBq/kg are administered and the average distance from the patient is 1 m; and the minimum mass specific activity (A_{1s}^{min}) needed to have $t_{probe}^{min} = 1$ s, These estimates are obtained assuming SUV=4, TNR=8.


Isotope	S_{res} (cps)	S_{HT} (cps)	R_{Far} (cps)	t_{probe}^{min} (s)	A_{1s}^{min} (MBq/kg)	DR_{med} ($\mu\text{Sv/hr}$)
^{18}F	38	35	33	>25	>75	38
^{31}Si	38	10	2.3	0.4	1.3	0.027
^{32}P	48	13	2.5	0.4	1.2	0.027
^{67}Cu	1.4	1.3	1.2	>25	>75	4.8
^{83}Br	9.6	1.8	0.3	1.2	3.7	0.29
^{90}Y	71	26	6.1	0.4	1.2	0.027
^{97}Zr	319	143	102	0.1	0.4	6.1
^{131}I	12	10	9.9	>25	>75	16
^{133}I	44	28	23	2.2	6.5	23
^{153}Sm	3.0	0.7	0.3	4.7	14	5.0
^{177}Lu	0.6	0.4	0.3	>25	>75	1.6
^{188}Re	55	18	4.3	0.5	1.4	2.1

Discussion

This study compares the performances of a prototype probe optimized for ^{90}Y detection on further radio-isotopes of interest. The results obtained are very important to guide on one side the choice of possible radio-tracers, on the other side the design of future prototypes.

The results indicate that three classes of radioisotopes can be identified:

- those that can be used with the current prototype alternatively to ^{90}Y : ^{31}Si , ^{32}P , ^{97}Zr , and ^{188}Re . Most of these isotopes are even more favorable than ^{90}Y since they have an electron energy endpoint low enough to limit the background from nearby healthy tissues, but high enough for the probe to be efficient. It is to be noted that lowering the minimum detectable energy would help anyhow, because it would allow detection also of partially infiltrated residues that are not completely at contact.



Acknowledgments

We would like to thank G. Chiodi and L. Recchia of the electronics shop of the INFN Sez. di Roma for the hardware and software support in the realization of the probe prototypes. This research was funded in part by AIRC-Fondazione Cariplo TRIDEO Id. 17515 and PRIN 2012 (prot. 2012CTAYSY).

Author Contributions

G.B., C.M.T., E.S.C., V.B., R.D., M.M., A.R., M.T. and S.M. designed and realized the experimental setups and took the data. C.M.T., F.C., R.F., E.S.C., and G.T. performed the simulations and the data analysis. A.C., I.F., D.R., and T.S. evaluated the relevance of the radio-isotopes from the chemical point of view while A.G. evaluated the potentialities from a Nuclear Medicine point of view. L.I. performed the radio-protection studies. Everybody contributed to the paper writing, coordinated by R.F.

Competing Financial Interests

F.C., M.M., and R.F. are listed as inventors on an Italian patent application (RM2013A000053) entitled “Utilizzo di radiazione β^- per la identificazione intraoperatoria di residui tumorali e la corrispondente sonda di rivelazione” dealing with the implementation of an intra-operative β^- probe for radio-guided surgery according to the results presented in this paper. The same authors are also inventors in the PCT patent application (PCT/IT2014/000025) entitled “Intraoperative detection of tumor residues using beta- radiation and corresponding probes ” covering the method and the instruments described in this paper.

References

- [1] Povoski, SP. et al. A comprehensive overview of radioguided surgery using gamma detection probe technology. *World Journal of Surgical Oncology* 7-11 (2009)
- [2] Tsuchimochi M. and Hayamaand K. Intraoperative gamma cameras for radioguided surgery: Technical characteristics, performance parameters, and clinical application. *Phys. Med.* **29**, 126-38 (2013)
- [3] Schneebaum S. et al. Clinical applications of gamma-detection probes - radioguided surgery. *Eur J Nucl Med.* 26: S26-35. (1999)
- [4] Solfaroli Camillocci E. et al. A novel radioguided surgery technique exploiting β^- decays. *Sci Rep.* 4: 4401 (2014)
- [5] Patent PCT/IT2014/000025 deposited by Università degli studi di Roma “La Sapienza”, Istituto Nazionale di Fisica Nucleare and Museo storico della fisica e centro studi e ricerche “E. Fermi”.

- [6] Mancini Terracciano M. et al. First Ex-Vivo Validation of a Radio-Guided Surgery Technique with β^- Radiation. arXiv:1608.08364. Accepted by Physica Medica (2016)
- [7] Bartolomei M. et al. Peptide receptor radionuclide therapy with ^{90}Y -DOTATOC in recurrent meningioma." *Eur J Nucl Med Mol Imaging* **36** 1407-16 (2009)
- [8] Collamati F. et al. Toward Radioguided Surgery with β^- Decays: Uptake of a Somatostatin Analogue, DOTATOC, in Meningioma and High-Grade Glioma. *J Nucl Med* **56** 1-6 (2015)
- [9] Collamati F. et al. Time evolution of DOTATOC uptake in Neuroendocrine Tumors in view of a possible application of Radio-guided Surgery with beta-Decays *J Nucl Med.* 56:1501-6 (2015)
- [10] for a review see Yeaong C.H. et al. Therapeutic radionuclides in nuclear medicine: current and future prospects *J. Zhejiang Univ. SCIENCE B Vol.15 No.10 P.845-863 (2014)* ;
- [11] Hickernell T. S. et al. Dual detector Probe for surgical Tumor Staging, *J. Nucl. Med.*, **29**, 1101 (1988); Daghighian, F. et al. Intraoperative beta probe: A device for detecting tissue labeled with positron or electron emitting isotopes during surgery *Med. Phys.* **21**, 153 (1994); Raylman, R. R. and Wahl R. L. A fiber-optically coupled positron-sensitive surgical probe, *J. Nucl. Med.*, **35**, 909 (1994); Bonzom S. An Intraoperative Beta Probe Dedicated to Glioma Surgery: Design and Feasibility Study, *IEEE Trans. Nucl. Sci.* **54**, 1 (2007).
- [12] Raylman R. R. et al. Fluorine-18-fluorodeoxyglucose-guided breast cancer surgery with a positron-sensitive probe: Validation in preclinical studies *J. Nucl. Med.*, **36**, 1869 (1995). Zervos E. E. et al. ^{18}F -Labeled Fluorodeoxyglucose Positron Emission Tomography-Guided Surgery for Recurrent Colorectal Cancer: A Feasibility Study *J. of Surg. Res.* **97**, 9 (2001). Bogalhas, F. et al. Development of a positron probe for localization and excision of brain tumors during surgery, *Phys. Med. Biol.* **54**, 4439 (2009). Gonzales S.J. et al. An analysis of the utility of handheld PET probes for the intraoperative localization of malignant tissue, *J. Gastrointest. Surg.* **15**, 358-55 (2011). Singh, B. et al. A hand-held beta imaging probe for FDG. *Ann. of Nucl. Med.* **27**, 203-208 (2013).
- [13] Piert M. et al. Positron detection for the intraoperative localisation of cancer deposits. *European Journal of Nuclear Medicine and Molecular Imaging*, **34(10)** 1534-1544 (2007); Franc B.L., Mari C., Johnson D., and Leong S.P. The role of a positron-and high-energy gamma photon probe in intraoperative localization of recurrent melanoma. *Clinical nuclear medicine*, **30(12)**:787-791 (2005); Essner R., Daghighian F., and Giuliano A.E. Advances in FDG PET probes in surgical oncology. *The Cancer Journal* **8(2)** 100-108 (2002).
- [14] Angelone M. et al. Properties of para-terphenyl as detector for alpha, beta and gamma radiation.' *IEEE Transactions on Nuclear Science* **61**, 3 1483-87 (2014)

- [15] Bocci V. et al. The ArduSiPM a compact trasportable Software/Hardware Data Acquisition system for SiPM detector. IEEE Nuclear Science Symposium and Medical Imaging Conference (NSS/MIC) DOI: 10.1109/NSS-MIC.2014.7431252 (2014)
- [16] Ferrari A. et al. FLUKA: a multi particle transport code. *Tech. Rep.* CERN-2005-10, INFN/TC05/11, SLAC-R-773 (2009)
- [17] Schneider W. et al. Correlation between CT numbers and tissue parameters needed for Monte Carlo simulations of clinical dose distributions. *Phys Med Biol* **45** 459-478 (2000)
- [18] Unger L.M. and Trubey D.K. Specific Gamma-Ray Dose Constants for Nuclides Important to Dosimetry and Radiological Assessment. ORLN/RSIC-45 (1982)
- [19] Kim Y.C. et al. Radiation Safety Issues in Y-90 Microsphere Selective Hepatic Radioembolization Therapy: Possible Radiation Exposure from the Patients. *Nucl Med Mol Imaging* **44(4)** 252260 (2010)
- [20] Russomando A. et al. An Intraoperative ?? Detecting Probe For Radio-Guided Surgery in Tumour Resection. arXiv:1511.02059 , accepted by IEEE-TNS (2015)